## **TECHNICAL DATA**

Model:	SILVER	CLASSIC/BASIC	COMPACT	COMPACT MICRO
Motor:	Induction type	Induction type	DC 24V	DC 24V
Motor speed:	500-1500 r.p.m.	1500 r.p.m.	3.100 r.p.m.	1.600-3.100 r.p.m.
Reduction:	1:6	1:6	1:8	1:8
Strokes/sec:	5/6/7/8	8.5 (50 Hz) 10 (60 Hz)	10	6/7/8/9/10/11
Setting:	keyboard/display	digital timer	digital timer	keyboard/display
Timer:	5/30/60/90/120	15-90sec	15-90sec	15-990 sec.
	150/180/210/240 sec			(15 by 15)
	continuous mode	continuous mode	continuous mode	continuous mode
Start / Stop:	door activated	door activated	door activated	door activated
Security:	double on door	double on door	double on door	double on door
Mains:	230 V, 50 Hz	230 V, 50 Hz	230 V, 50 Hz	230 V, 50 Hz
	110 V, 60 Hz	110 V, 60 Hz	110 V, 60 Hz	110 V, 60 Hz
Power:	270 W	270 W	50 W	55 W
Dimensions:	24x30x40 cm	24x26x40 cm	15x15x30 cm	15x15x30 cm
(WxHxD)	9.4x11.8x15.7 in.	9.4x10.2x15.7 in.	5.9x5.9x11.8 in.	5.9x5.9x11.8 in.
Weight:	20 Kg/44 lb.	17 Kg/37.5 lb.	7 Kg/15.4lb	7 Kg/15.4lb

Specifications subject to change without notice

## **ORDERING INFORMATION**

Model	Volume	Cat. No.
BASIC	400 ml	470
BASIC PANORAMIC	400 ml	475
COMPACT	80 ml	085
CLASSIC	400 ml	400
SILVER	400 ml	420
CLASSIC PANORAMIC	400 ml	500
COMPACT MICRO	80 ml	080
SILVER PANORAMIC	400 ml	520
the state of the		

Specify mains required (Cat.No. may change)

## OPTIONAL ACCESSORIES

PORTABAG stainless steel holder for 6 bags (400 ml)
PORTABAG stainless steel holder for 6 bags (80 ml)
Bags with filter and without are supplied indifferent presentations

All the instruments carry a 3-year warranty on the mechanical parts.

All systems come with 50 polyethylene bags, user's guide, and power cable.



## Contact us in internet.

To see how the instruments works in a multimedia demonstration, to buy instruments, accessories or consumables on-line visit our web page:







E C A


Empresa
Registrada
850/ER/05/03
ISO 9001

Cat no.

2300 2380



# **MASTICATOR**





## Bar reader

The bar reader device can be installed easily in both: the SILVER model and in the COMPACT MICRO. It can reads the labels prior to start the process.

Put the sample into the sterile bag

Select the operation time

When the door is closed the operation

begins

## The principle of the system

The two paddles gently, but thoroughly homogenize the sample with the diluent. The diluent penetrates inside the sample tissues and is pushed out again, creating a perfect solution for further analysis.







## **Inside the instrument**

The perfect design of the connecting rods combined with the progressive strength of springs make unnecessary the use of special bags. The microorganisms are forced to join the liquid phase.

Pick up the bag with the solution ready to be analized.



Via Umberto Giordano, 5 - 35132 Padova Tel 049 2021144 - Fax 049 2021143 www.zetalab.it - email: info@zetalab.it

